

DESCRIPTION

Product	RVP-1230L, Influenza A Reporter Virus Particles (RVPs)
Lot	IAL-1156A
Subtype	H1N1
Strain	Missouri/11/2025
Reporter	<i>Renilla</i> Luciferase
Size	1.0 mL/vial
Packaging	10% FBS/DMEM/10 mM HEPES
Recommended Starting Input	1.25 μ L per well (96-well plate) for a S:B $\geq$ 200*
Mycoplasma Test	Negative
Manufacture Date	July 2026
Expiration Date	July 2029

SAFETY & HANDLING

Shipping	Shipped on dry ice
Stability and Storage	Store at $\leq -80^{\circ}\text{C}$ upon receipt


Influenza A RVPs are used to test the ability of serum, antibodies, and drugs to neutralize Influenza A infectivity. RVPs display antigenically correct HA and NA proteins pseudotyped on replication-incompetent virus particles that contain a heterologous lentiviral (HIV) core. RVPs are capable of a single round of infection and carry a genome that expresses either a GFP or luciferase optical reporter gene upon infection.

Influenza A RVPs are produced in HEK-293T cells using four separate plasmids, encoding the HA protein, the NA protein, a lentiviral gag polyprotein, and a reporter gene. RVPs are created using a second-generation lentiviral system with components that are highly unlikely to recombine to produce a fully infectious virus (requiring 3 separate recombination events to do so). However, lentiviruses are capable of genomic integration and RVPs are derived from biological materials so should be handled with caution within a BSL2 or enhanced BSL2 laboratory environment. RVPs are not to be used in humans or in animals raised for food.

Thaw tubes in a 37°C water bath for 3 minutes and place on ice until ready to use. RVPs will appear as a translucent, pink solution. Gently mix prior to use and pulse tube for 3 seconds at high speed in a tabletop microfuge to recover all volume from the tube. Vortexing of RVPs should be avoided. Re-freezing of RVPs is not recommended.

* Determined using the MDCK-SIAT1 cell line. Input should be optimized based on specific assay conditions and experimental needs. Signal-to-background values may vary depending on instrument make, model, and/or calibration.

INFECTIVITY DATA




Infectivity determined in MDCK-SIAT1 cells. Infectivity data represents the average of three independent vials, each tested in technical quadruplicate.

Renilla luciferase activity measured using the Promega *Renilla* luciferase assay system (Promega #E2710). Sample luminescence was read using a PerkinElmer EnVision plate reader.

SIGNAL TO BACKGROUND	
RVP 384-well Input (μL)	Signal: Background
20	10,350
10	4,566
5	2,857
2.5	1,918
1.25	1,301
0.63	659
0.31	331

Signal to background is calculated using mock infected cells as the negative control value.



Christina Go, PhD

Senior Application Scientist, Virology


cgo@integralmolecular.com

Integral Molecular

One uCity Square

25 N. 38th Street, Suite 800

Philadelphia, PA 19104



Jennifer Kilgore

Senior Project Manager/Quality Assurance, Virology

jkilgore@integralmolecular.com

Integral Molecular

One uCity Square

25 N. 38th Street, Suite 800

Philadelphia, PA 19104